

Reaction to Fire Test - Ignitability of Building Products

SPECIFICATIONS

A device for the basic characteristics for classification to the fire reaction class B – F.

MEASURING METHODS

Ignitability test according to ČSN EN ISO 11925-2 Fire reaction tests - Ignitability of building products subjected to direct impingement of flame - Part 2: Small flame source test.



OVERVIEW OF MEASURABLE PARAMETERS / OUTPUT INFORMATION

To determine the fire reaction class B – F according to ČSN EN ISO 13501.

TYPES OF SAMPLES SUITABLE FOR ANALYSIS / MEASUREMENT CONDITIONS

Building products / materials including floor coverings measuring 250 x 90 mm.